Visualizing Cortical Response to Optogenetic Stimulation and Sensory Inputs Using Multispectral Handheld Optoacoustic Imaging

Saak V. Ovsepian, Yuanyuan Jiang, Thomas C.P. Sardella, Jaber Malekzadeh-Najafabadi, Neal C. Burton, Xin Yu, Vasilis Ntziachristos



PII:	S2213-5979(19)30076-X
DOI:	https://doi.org/10.1016/j.pacs.2019.100153
Reference:	PACS 100153
To appear in:	Photoacoustics
Received Date:	25 July 2019
Revised Date:	28 November 2019
Accepted Date:	5 December 2019

Please cite this article as: Ovsepian SV, Jiang Y, Sardella TCP, Malekzadeh-Najafabadi J, Burton NC, Yu X, Ntziachristos V, Visualizing Cortical Response to Optogenetic Stimulation and Sensory Inputs Using Multispectral Handheld Optoacoustic Imaging, *Photoacoustics* (2019), doi: https://doi.org/10.1016/j.pacs.2019.100153

This is a PDF file of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability, but it is not yet the definitive version of record. This version will undergo additional copyediting, typesetting and review before it is published in its final form, but we are providing this version to give early visibility of the article. Please note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

© 2019 Published by Elsevier.

Brief Communication

Visualizing Cortical Response to Optogenetic Stimulation and Sensory Inputs Using Multispectral Handheld Optoacoustic Imaging

Authors: Saak V. Ovsepian ^{1,2,3,4}, Yuanyuan Jiang ⁵, Thomas C.P. Sardella ⁶, Jaber Malekzadeh-Najafabadi ^{1,2}, Neal C. Burton ⁶, Xin Yu ^{5,7}, Vasilis Ntziachristos ^{1,2}

Affinitions: Institute for Biological and Medical Imaging, Helmholtz Zentrum Munich, German Research Center for Environmental Health, Neuherberg, Germany (1); Munich School of Bioengineering and Chair of Biological Imaging at Technical University Munich, Munich, Germany (2); Department of Experimental Neurobiology, National Institute of Mental Health, Klecany, Czech Republic (3); Department of Psychiatry and Medical Psychology, Third Faculty of Medicine, Charles University, Praha 10, Czech Republic (4); High-Field Magnetic Resonance Department, Max Planck Institute for Biological Cybernetics, Tuebingen, Germany (5); iThera Medical GmbH, Munich, Germany (6); Athinoula A. Martinos Center for Biomedical Imaging, Massachusetts General Hospital and Harvard Medical School, Charlestown, MA, USA (7).

Corresponding authors: Saak V. Ovsepian (saak.ovsepian@gmail.com) and Vasilis Ntziachristos (v.ntziachristos@helmholtz-muenchen.de).

ABSTRACT

To date, the vast majority of intra-vital neuroimaging systems applied in clinic and diagnostics is stationary with a rigid scanning element, requires specialized facilities and costly infrastructure. Here, we describe a simple yet radical approach for optoacoustic (photoacoustic) brain imaging *in vivo* using a light-weight handheld probe. It enables multispectral video-rate visualization of hemoglobin gradient changes in the cortex of adult rats induced by whisker and forelimb sensory inputs, as well as by optogenetic stimulation of intra-cortical connections. With superb penetration and molecular specificity, described here method holds major promises for future applications in research, routine ambulatory neuroimaging, and diagnostics.

Keywords: neuroimaging; hemoglobin gradients; brain; optogenetic stimulation; photoacoustic; handheld probe; barrel cortex

MAIN

Localized changes in perfusion and metabolic activity in the brain are closely linked to its functions and provide a contrast mechanism for imaging. Since the first demonstration of increased pulsation in the brain of subjects engaged in mathematical tasks [1], the field of functional neuroimaging has made drastic advances in basic and translational domains. As the largest consumer of glucose and oxygen, the mammalian and especially human brain absorbs excessively large amount of energy [2-4]. Over ~80% of this energy is used for maintaining baseline electrochemical processes,

with remaining ~20% accounts for processing sensory inputs and higher integrative functions and driving motor activity. While moderate, transitory changes in energy demands and fluctuations in tissue perfusion present a highly valuable source of information about physiological and pathological processes in the brain [5-7]. Based on blood-oxygenation level-dependent (BOLD) signal, for instance, major breakthroughs have been made by functional magnetic resonance imaging (fMRI) while the ability to sense small increments in metabolic activity renders positron emission tomography (PET) a powerful tool for molecular imaging [8, 9]. Although providing excellent amenities to neuroscience research and diagnostics, these methods fall short in yielding high-resolution data, come at poor scanning flexibility, very high costs and risks related to the use of radioactive tracers.

Optical imaging has been traditionally viewed as a low-cost alternative, yielding excellent spatial and temporal resolutions, but at the expense of limited penetration, due to scattering of the light by the brain tissue and skull [10, 11]. As an inherently hybrid method, optoacoustic (photoacoustic) imaging capitalizes on excellent resolution and molecular absorption of light as a contrast source, while emitted non-radiative decay broadband ultrasound (US) is used for image formation [12-14]. Such useful combination of superb contrast and molecular specificity of optical imaging with penetration and resolution of US render this method highly instructive for a wide range of biomedical applications. With added multiplexing and volumetric capabilities, recently, a multispectral optoacoustic tomography (MSOT) has been introduced, capable of imaging at unprecedented molecular specificity, resolution, and depth [14-17]. Unlike stationed fMRI and PET neuroimaging platforms, optoacoustic approaches offer excellent flexibility, with a portable scanning element applied for the research and diagnostics [18-20]. Up until now, however, neuroimaging applications of optoacoustic methods have been limited to the use of fixed detectors,

or their move with defined geometry [16, 21, 22]. Attempts of developing wearable detectors for imaging moving animals have been made recently, with sensors rigidly mounted on the animal head with permanent implants [23], an approach that has dubious clinical prospects.

We implemented for the first time a portable MSOT for sensing hemodynamic response induced by two independent sensory inputs and optogenetic stimulation of intra-cortical connections within the barrel field of adult anesthetized rats *in vivo*. Figure 1 (A) depicts a schematic diagram of the system with description of main components of the handheld probe, and anatomical image of the brain acquired at 720 nm, while (B) shows simulated sensitivity distribution field and detection band, and compares it with the size of the adult rat skull. Procedures on animal preparation were similar to those detailed elsewhere [24], and are provided in detail in Supplementary Materials and Methods.



Figure 1: (A) A schematic of the MSOT system with a portable handheld probe (top) employed for functional imaging of rat brain, with anatomical images captured at 720 nm without and with a superposition of the brain map at -0.5 mm Bregma (bottom, left and middle). Imaging planes containing S1FL and S1BF (-2.1 and 0.5 mm, respectively, bottom right) have been captured for current analysis. FB - fiber bundle, EC - electrical cables, FBO - fiber bundle output, TA - transducer array; IA - imaging aperture. (B) Simulated sensitivity distribution of the array within detection plane (left, middle) overlaid with the rat skull (right). TA – transducer array. (C) Fast Fourier transform graph of reconstructed data at 720 nm from the S1BF before and after stimulation of the whiskers input (1 and 2, respectively). Insets show frequency maps of imaging

planes superimposed with schematic anatomical representations, with color-coded deoxy-HB gradients under rest (baseline) and stimulation. (D) Normalized absorptivity graphs of four key absorbers in the brain within the wavelength range used for this study, with three wavelengths (dashed vertical lines) selected in current analysis.

To ensure optimal access to brain tissue, anesthetized with isoflurane rats were fixed in a stereotactic frame with two cranial windows (diameter = 5 mm) opened over the S1BF and S1FL fields [25], which were covered with medical Vaseline for protection, and shielded by transparent polyethylene (PE) foil. The acoustic gel was applied to ensure coupling between the brain surface and the detector membrane of the scanner. A custom-designed elastic handle was used for supporting the probe, to prevent excessive mechanical pressure on the brain tissue during measurements. After surgery, rats were given a bolus of α -chloralose (80 mg/kg i.v.) and isoflurane was discontinued. The infusion of α -chloralose was set at a rate of 26.5 mg/kg per hour. Pancuronium bromide (2 mg/kg/h i.v.) was delivered to achieve muscle relaxation and to minimize motion artifacts.

To stimulate the forepaw and whisker representations in the cortex, incrementing electrical pulses were delivered through bipolar electrodes to the forepaw and whisker area. Pulse sequencebased trigger and stimulation control were established using Bio-Pac system (MP 150 system, Goleta, USA). The baseline and stimulated activity responses were collected at 720 and 900 nm, which correspond to optimal absorption for oxyhemoglobin and deoxyhemoglobin (Deoxy-HB), while 800 nm was used for measurements of changes in the perfusion rate. Fast Fourier Transform (FFT) analysis was applied to calculate frequency components of collected signals before and after stimulation (Fig. 1C, D), with reconstructed hemoglobin gradient maps showing activated region by plotting frequency components corresponding to the maximal frequency distribution of each

pixel of raw images. As illustrated in Fig. 2 (A1-3, B1-3), electrical stimulation of the forepaw or whisker area by three consecutive trains of stimuli induced rapid onset and localized surge in deoxyhemoglobin in the front-limb and barrel field area of the contralateral somatosensory cortex (S1FL and S1BF). These transient changes in hemoglobin gradients were paralleled by remarkable increase in tissue oxygenation (SO2) within activated regions, which showed close temporal characteristics to deoxy-HB signals. As can be seen, evoked hemoglobin gradient and SO2 changes displayed a fast on- and off-rates locked in time with stimulation trains sequences (Fig. 2A3, B3). In total 68 trials in three independent rats, these evoked responses were highly reproducible and could be triggered by relatively low stimulus intensities. Imaging of anatomically matching ipsilateral cortex revealed no changes in hemoglobin gradients, confirming the specificity of evoked signals.

To find out if hemodynamic changes associated with cortical activation could be induced by optogenetic stimulation of intercortical circuits, ~1-month-old Sprague-Dawley rats were injected with vectors (AAV5.CaMKIIa.hChR2 (H134R)-eYFP) in S1BF as described previously [24]. After 4-6 weeks of recovery time and hChR2 expression in neurons, a fine optic fiber (diameter=0.2 mm) was inserted into the rat brain to stimulate cortical circuits using a 473 nm laser (CNI, China) with a built-in FC/PC coupler to deliver light pulses to the transfected area (Suppl. Video 1). As illustrated in Fig. 2C1-3, similar to electrical stimulation, activation of cortical circuits with light-induced a rapid hemodynamic response, which at low stimulus intensities was limited to the activation site, which was locked in time with light stimulation (Fig. 2C1-3). Stronger stimulations, however, induced a diffuse response spreading over the neighboring areas and extended in time beyond the stimulation period, suggesting induction of generalized seizure response.





Figure 2: (A1-C1) Schematic of the experimental setup used for recording the activation response in S1FL and S1BF by electrical stimulation of the forepaw and whisker pad area (A1 and B1), and by optogenetic stimulation of the S1BF (C1). (A2-C2) FFT maps of imaging planes superimposed with schematic of anatomical representation of corresponding planes, with color-coded deoxy-HB gradients shown under rest and stimulation. Note that for optical stimulation, the fiber was inserted directly into the barrel cortex of the stimulated side. S – Stimulation site. (A3-C3) Representative traces of deoxy-HB signals and calculated oxygen saturation (SO2) under rest and stimulation (top) with corresponding stimulation protocols (bottom). IL-ipsilateral; CL-contralateral.

Taken together, presented herein data demonstrate the feasibility of mapping hemodynamic changes related to processing somatosensory input in the cerebral cortex using a handheld MSOT.

To the best of our knowledge, this is the first report of multispectral functional neuroimaging using a hand-held optoacoustic probe, validating the possible use of portable brain imaging probes for research, potentially also extendable to the clinic and ambulatory settings. The principal advantage of our approach versus reported earlier wearable optoacoustic imaging is that it affords flexibility for a multispectral interrogation with excellent sensitivity, whereas, with wearable approach, only limited number of wavelengths is available. Also, the wearable optoacoustic imaging is more invasive as it involves removal of the skull with hard mounting of the transducer array to the skull. Finally, the portable approach presented here offers a wide range of potential translational applications in humans, particularly in infants where optical and acoustic properties of tissue afford an excellent window for mapping hemoglobin gradients and their accurate quantification. Given the superb penetration of optoacoustic interrogation with excellent contrast and specificity [15, 26, 27], further optimization of described approach is expected to facilitate routine use of handheld probe for brain MSOT. Detection of vascular and metabolic processes in human brain, as well as sensing hemorrhage, stroke, and brain tumors, are of particular interest given that they are associated with localized changes of hemoglobin gradients. Successful implementations of noninvasive optoacoustic neuroimaging in larger animals, including sheep and monkey [28, 29], and relatively lower costs render handheld MSOT attractive for future use in biomedical research and healthcare benefits.

Conflict of Interest

A conflicting interest exists when professional judgement concerning a primary interest (such as patient's welfare or the validity of research) may be influenced by a secondary interest (such as financial gain or personal rivalry). It may arise for the authors when they have financial interest that may influence their interpretation of their results or those of others. Examples of potential conflicts of interest include employment, consultancies, stock ownership, honoraria, paid expert testimony, patent applications/registrations, and grants or other funding.

Dr. Thomas C.P. Sardella and Dr. Neal C. Burton work as imaging and application specialists at iThera Medical GmbH.

Please state any sources of funding for your research

This work was supported by G.W. Leibnitz Prize DFG (2013) to V.N., PRISAR grant (EU) to V.N., Helmholtz Association Developmental Grant to S.V.O.

Funding Source

All sources of funding should also be acknowledged and you should declare any involvement of study sponsors in the study design; collection, analysis and interpretation of data; the writing of the manuscript; the decision to submit the manuscript for publication. If the study sponsors had no such involvement, this should be stated.

Acknowledgments: This work was supported by G.W. Leibnitz Prize DFG (2013) to V.N., PRISAR grant (EU) to V.N., Helmholtz Association Developmental Grant to S.V.O.

References

- 1. Mosso, M., Uber den Kreislauf des Blutes im Menschlichen Gehirn. Verlag von Veit und Company., 1881.
- Clark, D.D., D.D. Clark, and L. Sokoloff, *Basic Neurochemistry: Molecular, Cellular and Medical Aspect.*, ed. G.J. Siegel, et al. 1999, Philadelphia: Lippincott.
- 3. Ovsepian, S.V., *The dark matter of the brain*. Brain Struct Funct, 2019. **224**(3): p. 973-983.
- 4. Raichle, M.E., *Behind the scenes of functional brain imaging: a historical and physiological perspective.* Proc Natl Acad Sci U S A, 1998. **95**(3): p. 765-72.
- Bandettini, P.A., et al., *Time course EPI of human brain function during task activation*. Magn Reson Med, 1992. 25(2): p. 390-7.
- 6. Kwong, K.K., et al., *Dynamic magnetic resonance imaging of human brain activity during primary sensory stimulation*. Proc Natl Acad Sci U S A, 1992. **89**(12): p. 5675-9.

- 7. Ogawa, S., et al., Brain magnetic resonance imaging with contrast dependent on blood oxygenation. Proc Natl Acad Sci U S A, 1990. 87(24): p. 9868-72.
- Hooker, J.M. and R.E. Carson, *Human Positron Emission Tomography Neuroimaging*. Annu Rev Biomed Eng, 2019. 21: p. 551-581.
- 9. Logothetis, N.K., *What we can do and what we cannot do with fMRI*. Nature, 2008. **453**(7197): p. 869-878.
- 10. Kerr, J.N. and W. Denk, *Imaging in vivo: watching the brain in action*. Nat Rev Neurosci, 2008.
 9(3): p. 195-205.
- Ntziachristos, V., *Going deeper than microscopy: the optical imaging frontier in biology*. Nature Methods, 2010. 7(8): p. 603-614.
- 12. Ntziachristos, V., et al., *Looking and listening to light: the evolution of whole-body photonic imaging*. Nat Biotechnol, 2005. **23**(3): p. 313-20.
- Oraevsky, A., et al., Laser opto-acoustic imaging of early mucosal cancer: Feasibility studies of a new diagnostic modality in a hamster model of oral cancer. Gastrointestinal Endoscopy, 1999.
 49(4): p. Ab158-Ab158.
- 14. Taruttis, A. and V. Ntziachristos, *Advances in real-time multispectral optoacoustic imaging and its applications*. Nature Photonics, 2015. **9**(4): p. 219-227.
- 15. Ntziachristos, V. and D. Razansky, *Molecular imaging by means of multispectral optoacoustic tomography (MSOT)*. Chem Rev, 2010. **110**(5): p. 2783-94.
- Olefir, I., et al., Spatial and Spectral Mapping and Decomposition of Neural Dynamics and Organization of the Mouse Brain with Multispectral Optoacoustic Tomography. Cell Rep, 2019.
 26(10): p. 2833-2846 e3.
- 17. Ovsepian, S.V., et al., *Pushing the Boundaries of Neuroimaging with Optoacoustics*. Neuron, 2017.
 96(5): p. 966-988.
- Buehler, A., et al., *Real-time handheld multispectral optoacoustic imaging*. Opt Lett, 2013. 38(9):p. 1404-6.

- Diot, G., et al., *Multispectral Optoacoustic Tomography (MSOT) of Human Breast Cancer*. Clin Cancer Res, 2017. 23(22): p. 6912-6922.
- Taruttis, A., et al., Optoacoustic Imaging of Human Vasculature: Feasibility by Using a Handheld Probe. Radiology, 2016. 281(1): p. 256-63.
- 21. Burton, N.C., et al., *Multispectral opto-acoustic tomography (MSOT) of the brain and glioblastoma characterization*. Neuroimage, 2013. **65**: p. 522-8.
- 22. Ovsepian, S.V., I. Olefir, and V. Ntziachristos, *Advances in Optoacoustic Neurotomography of Animal Models*. Trends Biotechnol, 2019.
- 23. Tang, J., et al., Wearable 3-D Photoacoustic Tomography for Functional Brain Imaging in Behaving Rats. Sci Rep, 2016. 6: p. 25470.
- 24. Yu, X., et al., *Sensory and optogenetically driven single-vessel fMRI*. Nat Methods, 2016. **13**(4): p. 337-40.
- 25. Paxinos, G. and C. Watson, *The rat brain in stereotaxtic coordinates*. 2007, Burlington, MA: Elsevier.
- Wang, L.V. and J. Yao, A practical guide to photoacoustic tomography in the life sciences. Nat Methods, 2016. 13(8): p. 627-38.
- Weber, J., P.C. Beard, and S.E. Bohndiek, *Contrast agents for molecular photoacoustic imaging*. Nat Methods, 2016. 13(8): p. 639-50.
- 28. Petrov, I.Y., et al., *Optoacoustic monitoring of cerebral venous blood oxygenation though intact scalp in large animals*. Optics Express, 2012. **20**(4): p. 4159-4167.
- 29. Yang, X.M. and L.V. Wang, *Monkey brain cortex imaging by photoacoustic tomography*. Journal of Biomedical Optics, 2008. **13**(4).

Biographie



Saak Victor Ovsepian is an Armenian-Irish neuroscientist, Ph.D. and Professor of Neurobiology, and the Director of the Department of Experimental Neurobiology at the National Institute of Mental Health, Czech Republic. He is also an Adjunct Professor of Neurotherapeutics at the International Centre for Neurotherapeutics, Dublin City University, Republic of Ireland. Dr. Ovsepian studied medicine, biology, and philosophy at Omsk State University and State University of Yerevan. In 1999 he earned his M.Sc. degree in human neurophysiology from the State University of Yerevan and L. Orbeli Brain Institute of Armenian Academy of Sciences. In 2003 he completed his Ph.D. training in comparative and evolutionary neurophysiology at Sechenov Institute of Evolutionary Physiology and Biochemistry, St. Petersburg Russian Federation. Throughout his carrier, he worked in several areas of neurobiology research, with the main focus on molecular biology and synaptic physiology, neuromodulation and neurodegenerative disease. Between 2008-2013 he served as the founding director of Neuroimaging and Drug Screening Laboratories at the International Centre for Neurotherapeutics at Dublin City University. From 2013 to 2018 he was the head of the Electrophysiology at the German Centre for Neurodegenerative Disease Research in Munich, and served as a Senior Investigator at the Institute of Biological and Medical Imaging, Helmholtz Zentrum Munich, Germany. Ovsepian has authored over 70 research articles, book chapters, and reviews on synaptic biology, brain physiology and evolution, development of bio-therapeutics, and neuroimaging. In 20014, he proposed the hypothesis of exaptive origin of chemical synapses, suggesting that the core constituents of chemical synapse have evolved before and independently from neuronal evolution, and have been subsequently co-opted and neo-functionalized for their new role. In 2014 he put forward the homeostatic hypothesis of basal forebrain cholinergic system dubbed as 'drain of the brain', which assigns previously unrecognized homeostatic role to basal forebrain cortical cholinergic projections in clearance of amyloid beta peptide from innervation fields. In 2019 Ovsepian formulated the brain's dark matter hypothesis, suggesting a massive redundancy of neurons in the mammalian nervous system, maintained in dormant state, which stayed out of the reach of natural selection, and therefore accumulated over the evolutionary process. Ovsepian's current research focus is on elucidating molecular and cellular mechanisms of neurodegenerative

and neuropsychiatric disease. Development and application of advanced optical imaging tools and methods for structural and functional brain research in animal models is also part of his research activities.